

CERTIFICATION OF QUALITY

Pall Corporation certifies that the specified product meets or exceeds specifications as outlined in our currently published literature. This product is HPLC certified.

Product Number:	A P - 4 5 5 8
Description:	Acrodisc® Premium Syringe Filter, 25 mm
Pore Size:	0.45 µm
Lot Number:	2 1 8 2 7 8 2 1
Shipment Information:	
Manufacture Date:	July 2013
Materials of Construction:	
GxF Prefilter:	Binder-free Borosilicate Glass
Filter Media:	GH Polypro (GHP) membrane (hydrophilic polypropylene)
Housing:	Polypropylene
Specifications:	
Effective Filtration Area:	3.9 cm ²
Typical Hold-Up Volume (with air purge):	< 200 µL
Typical Water Flow Rate:	195 mL/min at 2.1 bar (210 kPa, 30 psi)
Maximum Operating Pressure:	4.1 bar (410 kPa, 60 psi) at 21 – 24 °C (70 – 75 °F)
Maximum Operating Temperature:	55 °C (131 °F) at 2.1 bar (210 kPa, 30 psi)

Authorized Signature:



Name: Jose Cubero

Title: Sr Product Quality Manager

Date: Tuesday, July 23, 2013