


Qingdao Haier Biomedical Co., LTD

Certificate of Analysis

Product name	10µl Transparent Low-Retention Filtered Tips, Boxed	Specifications	96Tips/Rack 50Racks/Case	Item NO.	111007	
Production date	2025.09.26	Batch number	26925025	Shipment quantity	2 boxes	
Expiring date	2028.09.25	Inspection date	2025.10.27	Sampling quantity	200pcs	
Serial number	Detection items	Inspection content / inspection basis		Test results		
1	Outward appearance	Black spots, bubbles, scratches, etc		Meet the requirements		
2	dimension	Conform to drawing specifications		Meet the requirements		
3	packing	Packing material, quantity correct, packing intact		Meet the requirements		
4	material	Meet the requirements of the drawing		Meet the requirements		
5	Endotoxin test	Not more than 0.1EU/ml		Endotoxin test		
6	Rnase test	There should be no DNase		Meet the requirements		
7	Dnase test	There should be no RNase		Meet the requirements		
8	airtightness	Sealed without leakage		Meet the requirements		
9	Load/Eject Velocity	uniform		Meet the requirements		
10	Resistant to high temperature and pressure	Resistant to high temperature and pressure		Meet the requirements		
11	Sterilization	Sterilization label		Meet the requirements		
12	Low adsorption	There is no residue. There are no walls		Meet the requirements		
Comprehensive judgments		OK				
Explanation: "Conclusion" can be: pass muster/qualified; Not Satisfiable/NA						
Add;No.280.Feng.Yuan.Road,High-Tech.Zone,Qingdao,China		Inspector:	TangXiaoYan		Reviewer:	LuoLiuJie
Phone: (86)400-6992017						
E-mail: liujie.luo@haierbiomedical.com						

Form number: KS-PZ-039 A/0

