


Qingdao Haier Biomedical Co., LTD

Certificate of Analysis

Product name	0.2ml PCR 8-strip tubes, transparent (with flat cap)	Specifications	125 Strips/Pack; 10 Packs/Case	Item NO.	222008
Production date	2025.10.25	Batch number	29825006	Lot quantity	10Cases
Expiring date	2028.10.24	Inspection date	2025.12.15	Sampling quantity	315pcs
Serial number	Detection items	Inspection content / inspection basis		Test results	
1	Outward appearance	Black spots, bubbles, scratches, etc		Meet the requirements	
2	dimension	Conform to drawing specifications		Meet the requirements	
3	packing	Packing material, quantity correct, packing intact		Meet the requirements	
4	material	Meet the requirements of the drawing		Meet the requirements	
5	Endotoxin test	Not more than 0.1EU/ml		Endotoxin test	
6	Rnase test	There should be no DNase		Meet the requirements	
7	Dnase test	There should be no RNase		Meet the requirements	
8	Airtightness	Sealed without leakage		Meet the requirements	
9	Membrane filtration sterilization	The filtered liquid must not contain any microorganisms		Meet the requirements	
10	Sterilization	Sterilization label		Meet the requirements	
Comprehensive judgments		OK			
Explanation: "Conclusion" can be: pass muster/qualified; Not Satisfiable/NA					
Add; No. 280. Feng. Yuan. Road, High-Tech. Zone, Qingdao, China		Inspector: TangXiaoYan		Reviewer: LuoLiuJie	
Tel : (86) 400-6992017					
Email : liujie.luo@haierbiomedical.com					

Form number: KS-PZ-039 A/0

