


Product Quality Inspection Report

Product name	15mlCentrifugeTube	Specifications	25Tubes/Pack 20Pack/Case	Item No.	311005
Production date	2025.03.07	Batch number	06525121	Shipment quantity	100 Case
Expiring date	2028.03.06	Inspection date	2025.04.16	Sampling quantity	500pcs
Serial number	Detection items	Inspection content / inspection basis		Test results	
1	Outward	Black spots, bubbles, scratches, etc		Meet the requirements	
2	dimension	Conform to drawing specifications		Meet the requirements	
3	packing	Packing material, quantity correct, packing intact		Meet the requirements	
4	material	Meet the requirements of the drawing		Meet the requirements	
5	Endotoxin test	Chinese pharmacopoeia (2020 edition)1143 bacterial endotoxin test <0.005EU/ML		Endotoxin test	
6	Rnase test	Specification for nuclease test		Meet the requirements	
7	Dnase test	Specification for nuclease test		Meet the requirements	
8	airtightness	Sealed without leakage		Meet the requirements	
9	Load/Eject Velocity	uniform		Meet the requirements	
10	Resistant to high temperature and pressure	Resistant to high temperature and pressure		Meet the requirements	
11	Sterilization	Sterilization label		Meet the requirements	
Comprehensive judgments					

Description: Check Results can be filled in: Meet the requirements, do not meet the requirements or fill in NA

Add;No.280.Feng.Yuan.Road,High-Tech.Zone,Qingdao,China	Inspector:	YaoChenChen	Reviewer:	LuoLiuJie
Tel : (86) 400-6992017				
E-mail: zhangjinxiu@haierbiomedical.com				

Form number: KS-PZ-039 A/0