


Product Quality Inspection Report

product name	15ml Centrifuge tube	REF	311005	Batch number	24524121
Production date	2024.09.02	Expiring date	2027.09.01	Inspection report date	2024.09.13
Packaging specifications	25 Tubes/Bag, 20Bags/Case	Shipment quantity	450 BOX	Sampling quantity	1250 pcs
Serial Number	Testing items	Inspection content/inspection basis			detection result
1	appearance	No obvious black spots, bubbles, pulling marks, etc			meet a requirement
2	size	Complies with drawing specifications			meet a requirement
3	package	The packaging materials and quantity are correct, and the packaging is intact			meet a requirement
4	Centrifuge testing	no leakage			meet a requirement
5	tightness	no leakage			meet a requirement
6	Easy to open and close	Smooth opening and closing without any impact			meet a requirement
7	High temperature and high pressure resistance	No deformation			meet a requirement
8	Endotoxin test	Chinese pharmacopoeia (2020 edition) 1143 bacterial endotoxin test <0.005EU/ML			<0.005EU/ML
9	Rnase detection	Nuclease Test Specification - No RNA			meet a requirement
10	Dnase detection	"Nuclease Test Specification" No DNA			meet a requirement
11	Sterilize	rradiation label should be red			meet a requirement
Comprehensive judgment					
Explanation: "Inspection results" can be filled in as: meet the requirements, do not meet the requirements, or NA					
Address: No. 280 Feng Yuan Road. High-tech Zone. Qingdao. 266109. P. R China		Inspector: Yao Chen Chen	Reviewer: Yang Li		
Phone: (86) 400-6992017					
E-mail: changrui.xiao@haierbiomedical.com					

Form number: KS-PZ-039 A/0

