


Qingdao Haier Biomedical Co.,Ltd.

Product Quality Inspection Report

|                     |                                                                                     |                 |                                                   |                       |           |
|---------------------|-------------------------------------------------------------------------------------|-----------------|---------------------------------------------------|-----------------------|-----------|
| Product name        | 125ml Erlenmeyer Flask                                                              | REF             | 642053                                            | Shipment quantity     | 1 BOX     |
| Production batch    | 25324132                                                                            | Production date | 2024/9/11                                         | Expiring date         | 2027/9/10 |
| Test items          | InspectionStandard                                                                  |                 | Test Result                                       | Result Judgement      |           |
| Exterior            | No black spots, bubbles, flashes, scratches, etc.                                   |                 | No black spots, bubbles, flashes, scratches, etc. | Meet the requirements |           |
| Structure size      | Meet the requirements of the drawings                                               |                 | Meet the requirements                             | Meet the requirements |           |
| Hydrophily          | Water vapor should condense into fish scales                                        |                 | Meet the requirements                             | Meet the requirements |           |
| Package             | Packing materials, quantity are correct, and packaging is intact                    |                 | Meet the requirements                             | Meet the requirements |           |
| Sterilize           | rradiation label should be red                                                      |                 | Irradiation label should is red                   | Meet the requirements |           |
| Bacterial endotoxin | Not more than 0.1EU/ml                                                              |                 | Not detected                                      | Meet the requirements |           |
| vitro cytotoxin     | Should meet the requirements of level 0                                             |                 | No cytotoxicity                                   | Meet the requirements |           |
| DNase detection     | There should be no DNase                                                            |                 | Not detected                                      | Meet the requirements |           |
| RNase detection     | There should be no RNase                                                            |                 | Not detected                                      | Meet the requirements |           |
| Cell culture        | Cells are in good condition and morphology                                          |                 | Cells are in good condition and morphology        | Meet the requirements |           |
| Conclusion          |  |                 |                                                   |                       |           |
| Explanation:        | "Conclusion" can be: pass muster/qualified; Not Satisfiable/NA                      |                 |                                                   |                       |           |

Inspector: YaoChenChen

Reviewer: YangLi

Approver: XiaoChangRui

Form number: KS-PZ-039 A/0

