

## Certificate of Quality

**REF**

651201

**LOT**

E260236B

Product description: MICROPLATE, 96 WELL, PP, V-BOTTOM ,  
(CHIMNEY WELL), NATURAL, 10 PCS./BAG

Following inspections were carried out and passed:

raw material	<input checked="" type="checkbox"/>	passed
visual	<input checked="" type="checkbox"/>	passed
physical	<input checked="" type="checkbox"/>	passed

Free of detectable DNase/RNase/human DNA  
Representative product was tested by Real-Time-PCR.  
Detection limits:  
DNase 4.54 × 10<sup>-9</sup> Kunitz units  
RNase 1.0 × 10<sup>-9</sup> Kunitz units  
hDNA 2.7 pg

Free of detectable endotoxins (pyrogens)  
Representative product was tested according to the United States Pharmacopoeia (USP) Chapter <85>  
Detection limit: 0.03 EU/ml

We are certified according to the Quality System Standard ISO 9001 and ISO 13485.

This certificate is valid for the above mentioned lot and has no validity if no lot number is provided. This document has been created digitally and is valid without a signature.



Certificate creation date: 2026-02-11

Page: 1/1

[www.gbo.com](http://www.gbo.com)