

## Certificate of Quality

**REF**

655090

**LOT**

E25123D4

Product description: CELL CULTURE MICROPLATE, 96 WELL, PS, F-BOTTOM (CHIMNEY WELL),  $\mu$ CLEAR®, BLACK, CELLSTAR®, TC, LID WITH CONDENSATION RINGS, STERILE, 8 PCS./BAG

Following inspections were carried out and passed:

raw material	<input checked="" type="checkbox"/>	passed
visual	<input checked="" type="checkbox"/>	passed
physical	<input checked="" type="checkbox"/>	passed

Free of detectable DNase/RNase/human DNA  
Representative product was tested by Real-Time-PCR.  
Detection limits:  
DNase  $4.54 \times 10^{-9}$  Kunitz units  
RNase  $1.0 \times 10^{-9}$  Kunitz units  
hDNA 2.7 pg

Free of detectable endotoxins (pyrogens)  
Representative product was tested according to the United States Pharmacopoeia (USP) Chapter <85>  
Detection limit: 0.03 EU/ml

Non-cytotoxic  
Representative product was tested according to ISO 10993-5 and has been shown to be non-cytotoxic.

We hereby certify that our products were manufactured and tested in accordance with the following international standards and meet the required specifications:

ISO 10993-5 "Biological evaluation of medical devices - Part 5: Tests for in vitro cytotoxicity"

ISO 11137 "Sterilization of health care products - Radiation"

We are certified according to the Quality System Standard ISO 9001 and ISO 13485.



2029-12-01

**STERILE**SAL  $10^{-3}$ 

This certificate is valid for the above mentioned lot and has no validity if no lot number is provided. This document has been created digitally and is valid without a signature.



Certificate creation date: 2026-02-25

Page: 1/1