

Certificate of Quality

REF

676001

LOT

20250973

Product description: EASYSEAL SEALER, CLEAR

Following inspections were carried out and passed:

Free of detectable DNase/RNase/human DNA
Representative product was tested by Real-Time-PCR.
Detection limits:
DNase 4.54×10^{-9} Kunitz units
RNase 1.0×10^{-9} Kunitz units
hDNA 2.7 pg

Free of detectable endotoxins (pyrogens)
Representative product was tested according to the United States Pharmacopoeia (USP) Chapter <85>
Detection limit: 0.03 EU/ml

We are certified according to the Quality System Standard ISO 9001 and ISO 13485.



2028-12-01

This certificate is valid for the above mentioned lot and has no validity if no lot number is provided. This document has been created digitally and is valid without a signature.



Certificate creation date: 2026-01-09

Page: 1/1

www.gbo.com