

## Certificate of Quality

**REF**

705071

**LOT**

E251236B

Product description: SINGLE-BREAK STRIP PLATE, 12XC8 WELL, PS, C-BOTTOM, CLEAR, MICROLON, HIGH BINDING, 10 PCS./BAG

Following inspections were carried out and passed:

raw material	<input checked="" type="checkbox"/>	passed
visual	<input checked="" type="checkbox"/>	passed
physical	<input checked="" type="checkbox"/>	passed

Free of detectable DNase/RNase/human DNA  
Representative product was tested by Real-Time-PCR.  
Detection limits:  
DNase 4.54 × 10<sup>-9</sup> Kunitz units  
RNase 1.0 × 10<sup>-9</sup> Kunitz units  
hDNA 2.7 pg

Free of detectable endotoxins (pyrogens)  
Representative product was tested according to the United States Pharmacopoeia (USP) Chapter <85>  
Detection limit: 0.03 EU/ml

Binding Property  
MICROLON® coefficient of variation < 5% (IgG ELISA / transmission measurement)  
FLUOTRAC# coefficient of variation < 10% (IgG ELISA / fluorescence measurement)  
LUMITRAC# coefficient of variation < 10% (IgG ELISA / luminescence measurement)

We are certified according to the Quality System Standard ISO 9001 and ISO 13485.



2029-11-23

This certificate is valid for the above mentioned lot and has no validity if no lot number is provided. This document has been created digitally and is valid without a signature.



Certificate creation date: 2026-02-17

Page: 1/1