

## Certificate of Quality

**REF**

762074

**LOT**

E250233W

Product description: STRIP PLATE, 12xF8, PS, F-BOTTOM,  
WHITE, LUMITRAC, HIGH BINDING, 10 PCS./BAG

Following inspections were carried out and passed:

raw material	<input checked="" type="checkbox"/>	passed
visual	<input checked="" type="checkbox"/>	passed
physical	<input checked="" type="checkbox"/>	passed

Free of detectable DNase/RNase/human DNA  
Representative product was tested by Real-Time-PCR.

Detection limits:

DNase 1.3 × 10<sup>-8</sup> Kunitz units  
RNase 5.0 × 10<sup>-11</sup> Kunitz units  
hDNA 5 pg

Free of detectable endotoxins (pyrogens)

Representative product was tested according to the United States Pharmacopoeia (USP) Chapter <85>  
Detection limit: 0.03 EU/ml

Binding Property

MICROLON® coefficient of variation < 5% (IgG ELISA / transmission measurement)  
FLUOTRAC# coefficient of variation < 10% (IgG ELISA / fluorescence measurement)  
LUMITRAC# coefficient of variation < 10% (IgG ELISA / luminescence measurement)

We are certified according to the Quality System Standard ISO 9001 and ISO 13485.



2029-01-19

This certificate is valid for the above mentioned lot and has no validity if no lot number is provided. This document has been created digitally and is valid without a signature.



Certificate creation date: 2025-03-03

Page: 1/1

[www.gbo.com](http://www.gbo.com)