

Certificate of Quality

REF

772350

LOT

A25113LV

Product description: Sapphire Pipette Tip, 10 µl XL,
natural, non-sterile,
1000 pcs./bag

Following inspections were carried out and passed:

raw material	<input checked="" type="checkbox"/>	passed
visual	<input checked="" type="checkbox"/>	passed
physical	<input checked="" type="checkbox"/>	passed

Non-cytotoxic

Representative product was tested according to ISO 10993-5 and has been shown to be non-cytotoxic.

Free of detectable DNase/RNase/human DNA/PCR-inhibitors

Representative product was tested by Real-Time-PCR and by an independent laboratory

Detection limits:

DNase 28.2 × 10⁻⁶ Kunitz units

RNase 1.6 fg

hDNA 32 pg

Free of detectable endotoxins (pyrogens)

Representative product was tested by LAL according to the European Pharmacopoeia (EP) 2.6.14.

Detection limit: <0.05 EU/ml

We hereby certify that our products were manufactured and tested in accordance with the following international standards and meet the required specifications:

ISO 10993-5 "Biological evaluation of medical devices - Part 5: Tests for in vitro cytotoxicity"

ISO 8655-6 "Piston-operated volumetric apparatus - Part 6: Gravimetric methods for the determination of measurement error"

ISO 9001 "Quality management systems - Requirements"



2028-10-26

This certificate is valid for the above mentioned lot and has no validity if no lot number is provided. This document has been created digitally and is valid without a signature.



Certificate creation date: 2025-12-01

Page: 1/1