

Certificate of Quality

REF

781096

LOT

E26043GB

Product description: MICROPLATE, 384 WELL, PS, F-BOTTOM, μ CLEAR®,
BLACK, MED. BINDING, 10 PCS./BAG

Following inspections were carried out and passed:

raw material	<input checked="" type="checkbox"/>	passed
visual	<input checked="" type="checkbox"/>	passed
physical	<input checked="" type="checkbox"/>	passed

Free of detectable DNase/RNase/human DNA
Representative product was tested by Real-Time-PCR.

Detection limits:

DNase 4.54×10^{-9} Kunitz units
RNase 1.0×10^{-9} Kunitz units
hDNA 2.7 pg

Free of detectable endotoxins (pyrogens)

Representative product was tested according to the United States Pharmacopoeia (USP) Chapter <85>
Detection limit: 0.03 EU/ml

Binding Property

MICROLON® coefficient of variation < 5% (IgG ELISA / transmission measurement)
FLUOTRAC# coefficient of variation < 10% (IgG ELISA / fluorescence measurement)
LUMITRAC# coefficient of variation < 10% (IgG ELISA / luminescence measurement)

We are certified according to the Quality System Standard ISO 9001 and ISO 13485.



2030-04-13

This certificate is valid for the above mentioned lot and has no validity if no lot number is provided. This document has been created digitally and is valid without a signature.



Certificate creation date: 2026-05-05

Page: 1/1

www.gbo.com